

FULLY AUTOMATED TOTAL THIOL ASSAY KIT

Short Name :TTL

Mod :End Point Monochromatic

Wavelength :405 nm

Unit : $\mu\text{mol/L}$

Decimal :3

R1 Volume :366 μl

R2 Volume :17 μl

Sample Volume: 17 μl

Incubation :8,5 min.

Kit Storage : 2-8 °C

Preparation : All reagents are ready to use

Equipment : Any Chemistry Analyser or Spectrophotometer

Method : Colorimetric

Exp Date : 6 months from the date of production

Handling : The product should be handled with the care usual when dealing with chemicals.

KIT COMPONENTS


Reagent 1 :50 ml

Reagent 2 :5 ml

Calibrator 2 :5 ml

CALIBRATION:

Two point



Calibrators	Materials	Value
Calibrator 1	Distiled Water	0 $\mu\text{mol/L}$
Calibrator 2	Inside the box	1000 $\mu\text{mol/L}$

Rel Assay Diagnostics Total Thiol Measurement Assay Kit

Intended Use

For the quantitative in vitro determination of total thiol content of serum/plasma and tissue/cell homogenates.

Introduction

Thiols are those compounds which contain the sulfhydryl group (-SH) attached to a carbon atom. They are endogenous molecules that assist aerobic cells maintaining a reducing state, despite an oxidizing environment. Thiol groups are found in all body cells and are indispensable for life. The detection and measurement of thiol content is one of the essential tasks for investigating biological processes and events in many biological systems.

Principle

Thiols interact with 5, 5'-dithiobis-(2-nitrobenzoic acid) (DTNB), forming a highly colored anion with maximum peak at 412 nm. The developed absorbance is proportional to total thiol content of the sample.

Specifications

- Colorimetric
- Manuel – spectrophotometric
- Fully Automated
- All reagents are liquid
- Ready to use
- Reagents are more stable and have long lifetimes
- Suitable for biological fluids, tissue/cell homogenates